

Solicitation 23164 - In-Breath Alcohol Testing Equipment and Accessories (Statewide) Bidder Question Responses

#	Solicitation Document Name	Document Section (Name or No.)	Comment/Question	Response
1	Solicitation	Section 2.4 - Lot 4 - Reference Standard (GAS)	In Lot 4, is the dry gas intended for use with the evidential Breath Analyzers, the Screening Devices or both?	Please see revised Section 3.4 - Lot 4 - Reference Standard (Gas) of the Solicitation (<i>Revised April 8, 2019</i>). The dry gas is intended for use for both evidential Breath Analyzers and the Screening Devices.
2	Solicitation	Section 2.4.1 - Lot 4A - Gas Reference Standard for Calibration of Breath Testing Equipment	It appears that the only differences between the sub-lots is the qualification of the bidder, as opposed to being any actual difference in the product itself. Please confirm.	Yes, per Section 2.41 - Lot 4A - Gas Reference Standard for Calibration of Breath Testing Equipment, the difference between the sub-lots is the qualification of the Bidder and there is no difference in the Product itself.
3	Solicitation	Section 3.1 - Lot 1 - Instrumentation	We intend to submit the Intoxilyzer 800 instrument which has been approved by NHTSA as an evidential breath test device and is listed among their current conforming products but is not listed on the older, published list that is found in section 3.1. We will provide the approval documentation with our bid submission or sooner if requested.	Please see revised Section 3.1 - Lot 1 - Instrumentation of the Solicitation (<i>Revised April 8, 2019</i>) which now states: Contract awards for the breath analyzers and screeners in Lot 1 will be limited to: (1) those products enumerated by the National Highway Traffic Safety Administration (NHTSA) in its published Conforming Products List (available at: https://www.govinfo.gov/content/pkg/FR-2017-11-02/pdf/2017-23869.pdf) or (2) those products which OGS can otherwise verify have been approved by NHTSA for inclusion on the Conforming Products List for instruments that conform to the Evidential Breath Testing Devices (82 FR 50940 and any subsequent NHTSA amending docket).
4	Attachment 1 - Pricing	Lot 3	The last Bid included spaces for pricing of solution in smaller Quantity lots. Is NY now going to purchase only in bulk and ship to only one location? Last years file is attached for your reference. As of March 2019 we are still supplying solutions in qtys of 2 bottles to many NY addresses.	An Authorized User may purchase Products in whatever unit size a Product is offered. Section 1.2 - Scope of the Solicitation (<i>Revised April 8, 2019</i>) states "A Bidder shall provide individual item pricing on any item or items on Attachment 1 - Pricing". In accordance with the "Instructions" for Attachment 1 - Pricing, Bidders are to enter the number of items in one unit for "unit size" on Attachment 1. Bidders may offer solutions in smaller quantity units.